

# Certificate



Certificate no.

CU 72181781 01

**License Holder:**

JustRight Surgical, LLC  
331 South 104th Street, Suite 200  
Louisville CO 80027  
USA

**Manufacturing Plant:**

Sparton Medical Systems  
Colorado LLC  
4300 Godding Hollow Parkway  
Frederick CO 80504  
USA

**Test report no.:** USA-AS 31290431 011

**Client Reference:** Claire Bronstein

**Tested to:** ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1  
IEC 60601-1-6:2010+A1  
IEC 60601-2-2:2009  
CAN/CSA-C22.2 No. 60601-1:14  
CAN/CSA-C22.2 NO. 60601-1-6:11 + A1  
CAN/CSA-C22.2 NO. 60601-2-2-09 (R2014)

**Certified Product:** RF Surgical Generator

**License Fee - Units**

**Model Designation:** JustRight Generator - JR100

7

**Rated Voltage:** AC 100-240V, 50-60Hz  
**Rated Current:** 1.5A  
**Protection Class:** I  
**Output Ratings:** 80Vrms, 250kHz, 25W at 250ohm

**Special Remarks:** Replaces Certificate CU72180103.

7

**Appendix:** 1, 1-4

**Licensed Test mark:**



**Date of Issue**  
**(day/mo/yr)**  
07/08/2018